
HELIOS®Plus-AMA A FAMILY

Albedo Sunlight Levels and Halogen Spectrums for Full Testing Solutions



Great solution for "Test As You Fly" applications

The HELIOSPlus-AMA, A Family systems are the ultimate expression of uniform calibration source technology allowing you to test with spectrums and levels that are equivalent to solar outputs. Labsphere has employed Xenon sources for bluer spectrum for many years, but in the A Family we have added cutting edge Electrode-less Plasma lamps that have incredible lifetime, light level outputs and radiometric stability. The A Family allows you to confirm all your testing with classic QTH and augment your characterization to test as you fly in with traceable solar-like light.

- Great solution for R&D and laboratory test flexibility
- Included CCD spectrometer and systems are compatible with Labsphere's full family of CDS spectrometer technology
- All options in HELIOSPlus available for A Family inclusion: Easy to configure a system to meet your exact requirements
- High level of absolute characterization

Performance

• Test As You Fly: Solar and QTH spectrums from 250 - 2500 nm

CREATE any spectrum

- Plasma 3000K to 5300K adjustable CCT and spectrum
- Xenon 3000K to 6500K adjustable CCT and spectrum
- Large Dynamic Light Range: >Albedo 0 (AM 0) down to night vision light levels
 - Cameras and sensors >16bit, actual 23bit (150dB)
 - Resolution and adjustability to meet customer's application and budget needs
- UV (<380 nm) ranges available



Model Number Smart Part Number	USLR-A20F-XAN2-P A5XA-N2NN-NNAR-NS00-0000-P	USLR-A20F-XDN2-P A5XD-N2NN-NNAR-NS00-0000-P	USLR-A20F-XMN2-P A5XM-N2NN-NNAR-NS00-0000-P	
OPTICAL PERFORMANCE SPECIFICATIONS UNIFORMITY (EACH LAMP OR COMBINED) Spatial Luminance Uniformity over Exit Port (f/4) - All Lamps On Angular Uniform FOV (Full Angle) - Degrees / F# / NA - All Lamps On	+/-1.0% +/-2.0% - 35° / 0.85 / 0.6	+/-1.0% +/-2.0% - 35° / 0.85 / 0.6	+/-1.0% +/-2.0% - 35° / 0.85 / 0.6	
QTH ONLY Expected Luminance Output: cd/m2 Expected Illuminance at Port: lux Est. Peak Radiance: W/m2-sr-um @ 0.95 um Est. Peak Irradiance @ Port: Photons/s-m2-um @ 0.95 um Minimum Resolution: lux Approximate Correlated Color Temperature (QTH) Typical Lamp Lifetimes (hrs) Est. Lamp Degradation Over Lifetime (% & CCT Shift) Est. Output Degradation over 50hrs (% & CCT Shift)	3,800 12,000 100 4.70E+20 1.20E-03 3000K +/-50K >500hrs -10% & +/-200K -1.0% & +/-20K	3,800 12,000 100 4.70E+20 1.09E+00 3000K +/-50K >500hrs -10% & +/-200K -1.0% & +/-20K	3,800 12,000 100 4.70E+20 2.40E-01 3000K +/-50K >500hrs -10% & +/-200K -1.0% & +/-20K	
PLASMA OR XENON ONLY Expected Luminance Output: cd/m2 Expected Illuminance at Port: lux Est. Peak Radiance: W/m2-sr-um @ 0.60um Pl @ 0.47 Xe Est. Peak Irradiance @ Port: Photons/s-m2-um @ 0.60um Pl or 0.47 Xe Minimum Resolution: lux Approximate Correlated Color Temperature Plasma Rest Mode/Xenon Flicker (Peak-Peak/RMS/Hz) - Typical Plasma Test Mode (Peak-Peak/RMS/Hz) - 30 minutes (Rest 5min) Installed Cold Mirror (330-750nm) Typical Lamp Lifetimes (hrs) Plasma/Xenon Est. Lamp Degradation Over Lifetime (% & CCT Shift) Est. Output Degradation over 50hrs (% & CCT Shift)	16,500 51,000 210 6.30E+20 5.10E-03 6000K +200K/-400K 4%/1%/5kHz N/A Yes 400 -50% & -600K -2.5% / -30K	16,500 51,000 210 6.30E+20 4.64E+00 6000K +200K/-400K 4%/1%/5kHz N/A Yes 400 -50% & -600K -2.5% / -30K	16,500 51,000 210 6.30E+20 1.02E-00 6000K +200K/-400K 4%/1%/5kHz N/A Yes 400 -50% & -600K -2.5% / -30K	
PLASMA/XENON & QTH (Both Full Open VA) Expected Luminance Output: cd/m2 Expected Illuminance at Port: lux Est. Peak Radiance: W/m2-sr-um @ 0.60um Pl @ 0.47 Xe Est. Peak Irradiance @ Port: Photons/s-m2-um @ 0.60um Pl @ 0.47 Xe Approximate Correlated Color Temperature (Xenon & QTH)	20,000 63,000 280 9.00E+20 5100K +400K/-200K	20,000 63,000 280 9.00E+20 5100K +400K/-200K	20,000 63,000 280 9.00E+20 5100K +400K/-200K	
ATTENUATORS Number of Steps in Attenuator Range Dynamic Range/Bits/dB - Full Range of System (both lamps)	2.00E+06 4.41E+07/25/152	1.20E+04 4.85E+04/15/93	1.00E+04 2.20E+05/17/106	
INTEGRATING SPHERE Coating / Material Sphere Internal Diameter: Inches (Meters) Frame Type Output Port Size: Inches (Meters)	Spectraflect® 20 (0.5) 20 in Cage 8 (0.2)	Spectraflect 20 (0.5) 20 in Cage 8 (0.2)	Spectraflect 20 (0.5) 20 in Cage 8 (0.2)	
SYSTEM COMPONENTS QTH Lamps Internal (#, Wattage) QTH Lamps External (#, Wattage) Xenon Lamp & Housing Plasma Lamp & Housing Power Supplies (# - Model) Special Power Supply Variable Attenuator Monitor Detector(s) Detector Filters (in Filter Holder) System Software Spectral Radiance Monitor (Type, Spectral Range)	(0) (1) 150 300W None (1) LPS-350 Integrated Xenon (2) VAA-220A SD-S1 Photopic HELIOSense CDS-610	(0) (1) 150 300W None (1) LPS-350 Integrated Xenon (2) VAD-012 SD-S1 Photopic HELIOSense CDS-610	(0) (1) 150 300W None (1) LPS-350 Integrated Xenon (2) VA-MM SD-S1 Photopic HELIOSense CDS-610	
STANDARD SYSTEM CALIBRATIONS (NIST Traceable) Luminance Correlated Color Temp (All lamps matched & w/VA position) QTH Only Spectral Radiance (350-2400nm) Xenon Only Spectral Radiance (350-2400nm) & CDS Exit Port Spatial Uniformity Exit Port Angular Uniformity Operational Duration of Calibration (Xe/QTH)	Yes Yes Yes Yes Yes 25 hrs / 50 hrs	Yes Yes Yes Yes Yes Yes 25 hrs / 50 hrs	Yes Yes Yes Yes Yes Yes 25 hrs / 50 hrs	

A runny. An Mass (Abcao) sunnyn	it source			TAUL Z
Model Number Smart Part Number	USLR-A12F-XAN2-P A3XA-N2NN-NNAR-NS00-0000-P	USLR-A12L-XAN2-P A4XA-N2NN-NNAR-NS00-0000-P	USLR-A12L-UAN1-P A4UA-N1NN-NNAR-NS00-0000-P	
OPTICAL PERFORMANCE SPECIFICATIONS UNIFORMITY (EACH LAMP OR COMBINED) Spatial Luminance Uniformity over Exit Port (f/4) - All Lamps On Angular Uniform FOV (Full Angle) - Degrees / F# / NA - All Lamps On	+/-1.0% +/-2.0% - 35° / 0.85 / 0.6	+/-1.0% +/-2.0% - 35° / 0.85 / 0.6	+/-1.0% +/-2.0% - 35° / 0.85 / 0.6	
QTH ONLY Expected Luminance Output: cd/m2 Expected Illuminance at Port: lux Est. Peak Radiance: W/m2-sr-um @ 0.95 um Est. Peak Irradiance @ Port: Photons/s-m2-um @ 0.95 um Minimum Resolution: lux Approximate Correlated Color Temperature (QTH) Typical Lamp Lifetimes (hrs) Est. Lamp Degradation Over Lifetime (% & CCT Shift) Est. Output Degradation over 50hrs (% & CCT Shift)	11,500 36,000 350 1.50E+21 3.60E-03 3000K +/-50K >500hrs -10% & +/-200K -1.0% & +/-20K	15,500 49,000 460 2.20E+21 3.60E-03 3000K +/-50K >500hrs -10% & +/-200K -1.0% & +/-20K	15,500 49,000 460 2.20E+21 3.60E-03 3000K +/-50K >500hrs -10% & +/-200K -1.0% & +/-20K	
PLASMA OR XENON ONLY Expected Luminance Output: cd/m2 Expected Illuminance at Port: lux Est. Peak Radiance: W/m2-sr-um @ 0.60um Pl @ 0.47 Xe Est. Peak Irradiance @ Port: Photons/s-m2-um @ 0.60um Pl or 0.47 Xe Minimum Resolution: lux Approximate Correlated Color Temperature Plasma Rest Mode/Xenon Flicker (Peak-Peak/RMS/Hz) - Typical Plasma Test Mode (Peak-Peak/RMS/Hz) - 30 minutes (Rest 5min) Installed Cold Mirror (330-750nm) Typical Lamp Lifetimes (hrs) Plasma/Xenon Est. Lamp Degradation Over Lifetime (% & CCT Shift) Est. Output Degradation over 50hrs (% & CCT Shift)	30,500 96,000 420 1.20E+20 9.60E-03 6000K +200K/-400K 4%/1%/5kHz N/A Yes 400 -50% & -600K -2.5% / -30K	42,000 132,000 550 1.60E+21 1.32E-02 6000K +200K/-400K 4%/1%/5kHz N/A Yes 400 -50% & -600K -2.5% / -30K	13,500 42,000 170 5.10E+20 4.20E-03 6000 +/-300K 4%/1%/5kHz N/A None 400 -50% & -600K -2.5% / -30K	
PLASMA/XENON & QTH (Both Full Open VA) Expected Luminance Output: cd/m2 Expected Illuminance at Port: lux Est. Peak Radiance: W/m2-sr-um @ 0.60um PI @ 0.47 Xe Est. Peak Irradiance @ Port: Photons/s-m2-um @ 0.60um PI @ 0.47 Xe Approximate Correlated Color Temperature (Xenon & QTH)	37,500 117,800 600 1.14E+21 5100K +400K/-200K	47,300 148,200 770 1.52E+21 5100K +400K/-200K	32,500 102,000 620 1.03E+21 5400K +400K/-200K	
ATTENUATORS Number of Steps in Attenuator Range Dynamic Range/Bits/dB - Full Range of System (both lamps)	2.00E+06 2.68E+07/24/148	2.00E+06 3.37E+07/25/150	2.00E+06 2.32E+07/24/146	
INTEGRATING SPHERE Coating / Material Sphere Internal Diameter: Inches (Meters) Frame Type Output Port Size: Inches (Meters)	Spectraflect 12 (0.3) 12 in Cage 4 (0.1)	Spectralon® 11.5 (0.29) 12 in Cage 4 (0.1)	Spectralon 11.5 (0.29) 12 in Cage 4 (0.1)	
SYSTEM COMPONENTS QTH Lamps Internal (# , Wattage) QTH Lamps External (#, Wattage) Xenon Lamp & Housing Plasma Lamp & Housing Power Supplies (# - Model) Special Power Supply Variable Attenuator Monitor Detector(s) Detector Filters (in Filter Holder) System Software Spectral Radiance Monitor (Type, Spectral Range)	(0) (1) 150 175W None (1) LPS-350 Integrated Xenon (2) VAA-220A SD-S1 Photopic HELIOSense CDS-610	(0) (1) 150 175W None (1) LPS-350 Integrated Xenon (2) VAA-220A SD-S1 Photopic HELIOSense CDS-610	(0) (1) 150 UV 175W None (1) LPS-350 Integrated Xenon (2) VAA-220A SD-S1 Photopic HELIOSense CDS-600	
STANDARD SYSTEM CALIBRATIONS (NIST Traceable) Luminance Correlated Color Temp (All lamps matched & w/VA position) QTH Only Spectral Radiance (350-2400nm) Xenon Only Spectral Radiance (350-2400nm) QTH & Xenon Spectral Radiance (350-2400nm) & CDS Exit Port Spatial Uniformity Exit Port Angular Uniformity Operational Duration of Calibration (Xe/QTH)	Yes Yes Yes Yes Yes Yes 25 hrs / 50 hrs	Yes Yes Yes Yes Yes Yes 25 hrs / 50 hrs	Yes Yes Yes Yes Yes Yes 25 hrs / 50 hrs	

A runny. An Muss (Albedo) sunny	it source			
Model Number Smart Part Number	USLR-A12F-XDN2-P A3XD-N2NN-NNAR-NS00-0000-P	USLR-A12L-XDN2-P A4XD-N2NN-NNAR-NS00-0000-P	USLR-A12L-UDN1-P A4UD-N1NN-NNAR-NS00-0000-P	
OPTICAL PERFORMANCE SPECIFICATIONS UNIFORMITY (EACH LAMP OR COMBINED) Spatial Luminance Uniformity over Exit Port (f/4) - All Lamps On Angular Uniform FOV (Full Angle) - Degrees / F# / NA - All Lamps On	+/-1.0% +/-2.0% - 35° / 0.85 / 0.6	+/-1.0% +/-2.0% - 35° / 0.85 / 0.6	+/-1.0% +/-2.0% - 35° / 0.85 / 0.6	
QTH ONLY Expected Luminance Output: cd/m2 Expected Illuminance at Port: lux Est. Peak Radiance: W/m2-sr-um @ 0.95 um Est. Peak Irradiance @ Port: Photons/s-m2-um @ 0.95 um Minimum Resolution: lux Approximate Correlated Color Temperature (QTH) Typical Lamp Lifetimes (hrs) Est. Lamp Degradation Over Lifetime (% & CCT Shift) Est. Output Degradation over 50hrs (% & CCT Shift)	11,500 36,000 350 1.50E+21 3.27E+00 3000K +/-50K >500hrs -10% & +/-200K -1.0% & +/-20K	15,500 49,000 460 2.20E+21 3.27E+00 3000K +/-50K >500hrs -10% & +/-200K -1.0% & +/-20K	15,500 49,000 460 2.20E+21 3.27E+00 3000K +/-50K >500hrs -10% & +/-200K -1.0% & +/-20K	
PLASMA OR XENON ONLY Expected Luminance Output: cd/m2 Expected Illuminance at Port: lux Est. Peak Radiance: W/m2-sr-um @ 0.60um Pl @ 0.47 Xe Est. Peak Irradiance @ Port: Photons/s-m2-um @ 0.60um Pl or 0.47 Xe Minimum Resolution: lux Approximate Correlated Color Temperature Plasma Rest Mode/Xenon Flicker (Peak-Peak/RMS/Hz) - Typical Plasma Test Mode (Peak-Peak/RMS/Hz) - 30 minutes (Rest 5min) Installed Cold Mirror (330-750nm) Typical Lamp Lifetimes (hrs) Plasma/Xenon Est. Lamp Degradation Over Lifetime (% & CCT Shift) Est. Output Degradation over 50hrs (% & CCT Shift)	30,500 48,000 420 1.20E+21 4.36E+00 6000K +200K/-400K 4%/1%/5kHz N/A Yes 400 -50% & -600K -2.5% / -30K	42,000 132,000 550 1.60E+21 1.20E+01 6000K +200K/-400K 4%/1%/5kHz N/A Yes 400 -50% & -600K -2.5% / -30K	16,500 51,800 430 7.16E+20 4.71E+00 6000 +/-300K 4%/1%/5kHz N/A None 400 -50% & -600K -2.5% / -30K	
PLASMA/XENON & QTH (Both Full Open VA) Expected Luminance Output: cd/m2 Expected Illuminance at Port: lux Est. Peak Radiance: W/m2-sr-um @ 0.60um PI @ 0.47 Xe Est. Peak Irradiance @ Port: Photons/s-m2-um @ 0.60um PI @ 0.47 Xe Approximate Correlated Color Temperature (Xenon & QTH)	42,000 84,000 600 1.43E+21 5100K +400K/-200K	58,000 182,000 820 2.50E+21 5100K +400K/-200K	32,000 100,800 620 1.03E+21 5400K +400K/-200K	
ATTENUATORS Number of Steps in Attenuator Range Dynamic Range/Bits/dB - Full Range of System (both lamps)	1.20E+04 2.95E+04/14/89	1.20E+04 3.71E+04/15/91	1.20E+04 2.55E+04/14/88	
INTEGRATING SPHERE Coating / Material Sphere Internal Diameter: Inches (Meters) Frame Type Output Port Size: Inches (Meters)	Spectraflect 12 (0.3) 12 in Cage 4 (0.1)	Spectralon 11.5 (0.29) 12 in Cage 4 (0.1)	Spectralon 11.5 (0.29) 12 in Cage 4 (0.1)	
SYSTEM COMPONENTS QTH Lamps Internal (#, Wattage) QTH Lamps External (#, Wattage) Xenon Lamp & Housing Plasma Lamp & Housing Power Supplies (# - Model) Special Power Supply Variable Attenuator Monitor Detector(s) Detector Filters (in Filter Holder) System Software Spectral Radiance Monitor (Type, Spectral Range)	(0) (1) 150 175W None (1) LPS-350 Integrated Xenon (2) VAD-012 SD-S1 Photopic HELIOSense CDS-610	(0) (1) 150 UV 175W None (1) LPS-350 Integrated Xenon (2) VAD-012 SD-S1 Photopic HELIOSense CDS-610	(0) (1) 150 UV 175W None (1) LPS-350 Integrated Xenon (2) VAD-012 SD-S1 Photopic HELIOSense CDS-600	
STANDARD SYSTEM CALIBRATIONS (NIST Traceable) Luminance Correlated Color Temp (All lamps matched & w/VA position) QTH Only Spectral Radiance (350-2400nm) Xenon Only Spectral Radiance (350-2400nm) QTH & Xenon Spectral Radiance (350-2400nm) & CDS Exit Port Spatial Uniformity Exit Port Angular Uniformity Operational Duration of Calibration (Xe/QTH)	Yes Yes Yes Yes Yes Yes 25 hrs / 50 hrs	Yes Yes Yes Yes Yes 25 hrs / 50 hrs	Yes Yes Yes Yes Yes Yes 25 hrs / 50 hrs	

, in mass (miscae) samigh				
Model Number Smart Part Number	USLR-A12F-XMN2-P A3XM-N2NN-NNAR-NS00-0000-P	USLR-A12L-XMN2-P A4XM-N2NN-NNAR-NS00-0000-P	USLR-A12L-UMN1-P A4UM-N1NN-NNAR-NS00-0000-P	
OPTICAL PERFORMANCE SPECIFICATIONS UNIFORMITY (EACH LAMP OR COMBINED) Spatial Luminance Uniformity over Exit Port (f/4) - All Lamps On Angular Uniform FOV (Full Angle) - Degrees / F# / NA - All Lamps On	+/-1.0% +/-2.0% - 35° / 0.85 / 0.6	+/-1.0% +/-2.0% - 35° / 0.85 / 0.6	+/-1.0% +/-2.0% - 35° / 0.85 / 0.6	
QTH ONLY Expected Luminance Output: cd/m2 Expected Illuminance at Port: lux Est. Peak Radiance: W/m2-sr-um @ 0.95 um Est. Peak Irradiance @ Port: Photons/s-m2-um @ 0.95um Minimum Resolution: lux Approximate Correlated Color Temperature (QTH) Typical Lamp Lifetimes (hrs) Est. Lamp Degradation Over Lifetime (% & CCT Shift) Est. Output Degradation over 50hrs (% & CCT Shift)	11,500 36,000 350 1.50E+21 7.20E-01 3000K +/-50K >500hrs -10% & +/-200K -1.0% & +/-20K	15,500 49,000 460 2.20E+21 9.80E+01 3000K +/-50K >500hrs -10% & +/-200K -1.0% & +/-20K	15,500 49,000 460 2.20E+21 9.80E+01 3000K +/-50K >500hrs -10% & +/-200K -1.0% & +/-20K	
PLASMA OR XENON ONLY Expected Luminance Output: cd/m2 Expected Illuminance at Port: lux Est. Peak Radiance: W/m2-sr-um @ 0.60um Pl @ 0.47 Xe Est. Peak Irradiance @ Port: Photons/s-m2-um @ 0.60um Pl or 0.47 Xe Minimum Resolution: lux Approximate Correlated Color Temperature Plasma Rest Mode/Xenon Flicker (Peak-Peak/RMS/Hz) - Typical Plasma Test Mode (Peak-Peak/RMS/Hz) - 30 minutes (Rest 5min) Installed Cold Mirror (330-750nm) Typical Lamp Lifetimes (hrs) Plasma/Xenon Est. Lamp Degradation Over Lifetime (% & CCT Shift) Est. Output Degradation over 50hrs (% & CCT Shift)	30,500 42,000 420 1.20E+21 8.40E+01 6000K +200K/-400K 4%/1%/5kHz N/A Yes 400 -50% & -600K -2.5% / -30K	42,000 132,000 550 1.60E+21 2.64E+00 6000K +200K/-400K 4%/1%/5kHz N/A Yes 400 -50% & -600K -2.5% / -30K	16,500 51,800 430 7.16E+20 1.04E+00 6000 +/-300K 4%/1%/5kHz N/A None 400 -50% & -600K -2.5% / -30K	
PLASMA/XENON & QTH (Both Full Open VA) Expected Luminance Output: cd/m2 Expected Illuminance at Port: lux Est. Peak Radiance: W/m2-sr-um @ 0.60um Pl @ 0.47 Xe Est. Peak Irradiance @ Port: Photons/s-m2-um @ 0.60um Pl @ 0.47 Xe Approximate Correlated Color Temperature (Xenon & QTH)	42,000 78,000 600 1.43E+21 5100K +400K/-200K	58,000 182,000 820 2.50E+21 5100K +400K/-200K	32,000 100,800 620 1.03E+21 5400K +400K/-200K	
ATTENUATORS Number of Steps in Attenuator Range Dynamic Range/Bits/dB - Full Range of System (both lamps)	1.00E+04 1.34E+05/17/102	1.00E+04 1.48E+05/17/102	1.00E+04 1.02E+05/16/99	
INTEGRATING SPHERE Coating / Material Sphere Internal Diameter: Inches (Meters) Frame Type Output Port Size: Inches (Meters)	Spectraflect 12 (0.3) 12 in Cage 4 (0.1)	Spectralon 11.5 (0.29) 12 in Cage 4 (0.1)	Spectralon 11.5 (0.29) 12 in Cage 4 (0.1)	
SYSTEM COMPONENTS QTH Lamps Internal (# , Wattage) QTH Lamps External (#, Wattage) Xenon Lamp & Housing Plasma Lamp & Housing Power Supplies (# - Model) Special Power Supply Variable Attenuator Monitor Detector(s) Detector Filters (in Filter Holder) System Software Spectral Radiance Monitor (Type, Spectral Range)	(0) (1) 150 175W None (1) LPS-350 Integrated Xenon (2) VA-MM SD-S1 Photopic HELIOSense CDS-610	(0) (1) 150 175W None (1) LPS-350 Integrated Xenon (2) VA-MM SD-S1 Photopic HELIOSense CDS-610	(0) (1) 150 UV 175W None (1) LPS-350 Integrated Xenon (2) VA-MM SD-S1 Photopic HELIOSense CDS-600	
STANDARD SYSTEM CALIBRATIONS (NIST Traceable) Luminance Correlated Color Temp (All lamps matched & w/VA position) QTH Only Spectral Radiance (350-2400nm) Xenon Only Spectral Radiance (350-2400nm) QTH & Xenon Spectral Radiance (350-2400nm) & CDS Exit Port Spatial Uniformity Exit Port Angular Uniformity Operational Duration of Calibration (Xe/QTH)	Yes Yes Yes Yes Yes Yes 25 hrs / 50 hrs	Yes Yes Yes Yes Yes Yes 25 hrs / 50 hrs	Yes Yes Yes Yes Yes Yes 25 hrs / 50 hrs	